

Hagen: Textbook of Diagnostic Sonography, 7th Edition

Chapter 01: Foundations of Sonography

Test Bank

MULTIPLE CHOICE

1. Historically, the development of ultrasound began shortly after:

- a. Radio communication in World War I
- b. Sonar in World War II
- c. Nuclear testing in World War II
- d. The launching of Sputnik

ANS: B

World War II brought sonar equipment to the forefront for defense purposes.

Ultrasound was influenced by the success of sonar equipment.

PTS: 1 REF: p. 5

OBJ: Detail a timeline for pioneers in the advancement of medical diagnostic ultrasound.

TOP: History of ultrasound

2. The early applications of obstetric ultrasound were initiated by:

- a. Joseph Holmes
- b. Ian Donald
- c. John Howry
- d. William Fry

ANS: B

The early obstetric compound scanner was built by Tom Brown and Dr. Ian Donald in Scotland in 1957.

PTS: 1 REF: p. 6

OBJ: Detail a timeline for pioneers in the advancement of medical diagnostic ultrasound.

TOP: History of ultrasound

3. Visualization of the cardiac structures in the heart was discovered by:

- a. Joseph Holmes
- b. Ian Donald
- c. Hertz and Edler
- d. George Ludwig

ANS: C

In 1954, echocardiographic techniques were developed in Sweden by Drs. C.H. Hertz and I. Edler.

PTS: 1 REF: p. 6

OBJ: Detail a timeline for pioneers in the advancement of medical diagnostic ultrasound.

TOP: History of ultrasound

4. Which one of the following statements about the role of sonographers is *false*?
- a. Sonographers perform ultrasound studies and gather diagnostic data independent of the physician.
 - b. Sonographers must possess intellectual curiosity and perseverance.
 - c. Sonographers must have a technical aptitude.
 - d. Sonographers must be able to communicate on different levels.

ANS: A

A sonographer performs ultrasound studies gathering diagnostic data under *both* the direct and the indirect supervision of a physician. They also must assess clinical history and symptoms, interpret laboratory values, and understand other diagnostic examinations.

PTS: 1 REF: p. 3

OBJ: Describe a career in ultrasound.

TOP: Role of the sonographer

5. In soft tissues, the assumed propagation velocity is (*in meters per second*):
- a. 1320
 - b. 1450
 - c. 1540
 - d. 1650

ANS: C

In soft tissues, the assumed propagation velocity (speed) is 1540 meters per second.

PTS: 1 REF: p. 8

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation

6. Diagnostic ultrasound uses the frequencies of:
- a. 10 to 15 kHz
 - b. 1 to 20 kHz
 - c. 100 to 1000 Hz
 - d. 1 to 20 MHz

ANS: D

Diagnostic application of ultrasound uses frequencies 1 to 20 million cycles per second (1 to 20 MHz).

PTS: 1 REF: p. 7

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation

7. Which one of the following correctly defines dynamic range?

- a. Field farthest from the transducer during formation of the sound beam
- b. Ratio of the largest to the smallest signals that an instrument can respond to without distortion
- c. Minimum distance between two structures positioned along the axis of the sound beam
- d. Measure of the strength of the ultrasound signal

ANS: B

Dynamic range is the ratio of the largest to the smallest signals that an instrument or component of an instrument can respond to without distortion.

PTS: 1 REF: p. 17

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation

8. The device that converts energy from one form to another is called the:

- a. Digitizer
- b. Transducer
- c. Scan converter
- d. Beam former

ANS: B

Piezoelectric elements (transducers) convert electric energy into ultrasound energy and vice versa.

PTS: 1 REF: p. 11

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation

9. The _____ is the product of the velocity of sound in a medium and the density of that medium.

- a. Acoustic impedance
- b. Sound reflection
- c. Angle of reflection
- d. Piezoelectric effect

ANS: A

The acoustic impedance is equal to the density multiplied by the sound propagation speed.

PTS: 1 REF: p. 9

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation

10. The angle of reflection is equal to the:

- a. Acoustic impedance
- b. Angle of incidence
- c. Refraction
- d. Image resolution

ANS: B

Angle of reflection is the angle between the reflected sound direction and a line perpendicular to the media boundary.

PTS: 1 REF: p. 5

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation

11. The best transducer to image the thyroid gland is:

- a. Curved array, 3 MHz
- b. Linear array, 5 MHz
- c. Sector array, 4 MHz
- d. Linear array, 3 MHz

ANS: B

High-frequency linear array transducers are generally used for smaller structures.

PTS: 1 REF: p. 12

OBJ: Identify ultrasound instruments and discuss their uses.

TOP: Sonography principles and instrumentation

12. The display mode that shows time along the horizontal axis and depth along the vertical axis is:

- a. A mode
- b. B mode
- c. M mode
- d. Real time

ANS: C

Motion mode (M mode) displays the depth along the vertical axis versus the time along the horizontal axis.

PTS: 1 REF: p. 14

OBJ: Identify ultrasound instruments and discuss their uses.

TOP: Sonography principles and instrumentation

13. Which one of the following statements about the Doppler principle is *false*?

- a. Doppler refers to a change in frequency in which the motion of laminar flow is detected within a vascular structure.
- b. The beam should be perpendicular to the flow.
- c. The Doppler shift is directly proportional to the velocity of the red cell.
- d. If the cell moves away from the transducer, then the fall in frequency is directly proportional to the velocity and direction of the red blood cell movement.

ANS: B

The beam should be parallel to the flow to obtain the maximum velocity. The frequency of the Doppler shift is proportional to the cosine of the Doppler angle. At a 90-degree angle (perpendicular to flow), the Doppler shift is zero, regardless of the flow velocity.

PTS: 1 REF: p. 17

OBJ: Discuss three-dimensional and Doppler ultrasound.

TOP: Sonography principles and instrumentation

14. The Fresnel zone is also called the:

- a. Far field
- b. Focal point
- c. Near zone
- d. Nyquist limit

ANS: C

The Fresnel or near zone is the field closest to the transducer during the formation of the sound beam.

PTS: 1 REF: p. 16

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation

15. The higher the transducer frequency, the:

- a. Shorter the wavelength
- b. Faster the frame rate
- c. Deeper the penetration depth
- d. Slower the frame rate

ANS: A

The higher the frequency, the shorter the wavelength.

PTS: 1 REF: p. 9

OBJ: Demonstrate an understanding of the basic principles and terminology of ultrasound.

TOP: Sonography principles and instrumentation